



9411 Corsair Road  
Frankfort IL 60423  
(800) 552-0299 Phone  
(815) 464-5650 Fax

**CERTIFICATE OF COMPLIANCE  
USP BIOLOGICAL TEST  
CLASSIFICATION USP CLASS VI**

The MG Series has been evaluated by NamSA and meets or exceeds the following requirements:

Acute Systemic Toxicity (USP) "T12"  
Intracutaneous Toxicity (USP) "T13"  
Implantation Test (USP) "14"

The test article(s) were evaluated following the guidelines of USP Class VI. In all cases the test article(s) met the requirements of the specification.

Certified By:   
Gary Johnson

Title: UP OPERATIONS

Date: 5/22/13

**NON-WARRANTY:** Information contained herein is based on tests we believe to be reliable and accurate. It is offered in good faith for the benefit of the consumer. Adhesive Systems shall not be liable for any injury, loss, or damage in the use of its chemical products since the conditions of use are beyond our control. In every case we urge and recommend the user conduct tests to determine to their own satisfaction that the product is of acceptable quality and is suitable for their particular purpose under their own operating conditions. Statements concerning the possible use of our products are not intended as recommendations to use our products in the infringement of any patent. Because of changing reporting requirements and other variables it is impossible to guarantee the accuracy of the information contained in this document. It is the responsibility of the user to determine proper personal protection based on the actual condition of use and to comply with all Federal, State, and Local laws and regulations.