



9411 Corsair Road  
Frankfort IL 60423  
(800) 552-0299 Phone  
(815) 464-5650 Fax

**CERTIFICATE OF CONFORMANCE  
ISO 10993-5  
CYTOTOXICITY STUDY**

**The following products meet the requirements of ISO 10993-5:**

MP531101M  
MP531102M  
MP531103M  
MP531104M  
MP431105M  
MP531106M  
MP531107M  
MP531108M  
MP531109M  
MP531110M  
MP531701M  
MP532109M  
MP532110M  
MP532111M  
MP532112M

**The test article(s) were evaluated for potential cytotoxic effects following the guidelines of ISO 10993-5, Biological evaluation of medical devices-Part 5: Tests for *in vitro* cytotoxicity. In all cases the test articles indicated no to slight toxicity. The test article(s) extract met the requirements of the test since grade in all cases was less than 2 (mild reactivity)**

**Certified By:**   
Gary Johnson

**Title:** VP OPERATIONS

**Date:** 5/22/13

**NON-WARRANTY:** Information contained herein is based on tests we believe to be reliable and accurate. It is offered in good faith for the benefit of the consumer. Adhesive Systems shall not be liable for any injury, loss, or damage in the use of its chemical products since the conditions of use are beyond our control. In every case we urge and recommend the user conduct tests to determine to their own satisfaction that the product is of acceptable quality and is suitable for their particular purpose under their own operating conditions. Statements concerning the possible use of our products are not intended as recommendations to use our products in the infringement of any patent. Because of changing reporting requirements and other variables it is impossible to guarantee the accuracy of the information contained in this document. It is the responsibility of the user to determine proper personal protection based on the actual condition of use and to comply with all Federal, State, and Local laws and regulations.